

## The Model T802 Paramagnetic O<sub>2</sub> Analyzer with Optional CO<sub>2</sub> Sensor



The Model T802  $O_2$  analyzer uses the proven paramagnetic measurement principle, coupled with state of the art microprocessor technology to provide accurate and dependable measurements of Percent Level  $O_2$ . The T802 offers a stable and inherently linear measurement of oxygen, making it possible to calibrate the analyzer by checking only two points. There is no requirement for a reference gas during operation.

— with NumaView<sup>™</sup> premium T Series software —

- Optional CO<sub>2</sub> Sensor
- Large, vivid, and durable color touchscreen display
- Lifetime technical support by phone and email
- All other T Series instrument platform features
- Standard two-year warranty





## Model T802 Specifications

Ranges	Min: 0-1% full scale Max: 0-100% full scale (selectable)
Zero Noise	< 0.02% (RMS)
Span Noise	< 0.05% of reading (RMS)
Lower Detectable Limit	< 0.04%
Zero Drift	< ±0.02% /24 hours; < ±0.05% /7 days
Span Drift	< ±0.1% /7 days
Accuracy	< ±0.1%
Temperature Coefficient	< ±0.01% per degree C
Linearity	< ±0.1%
Rise and Fall Time	< 60 seconds to 95%
Flow Rate	120 ml ± 20ml/min
Humidity Range	0 - 95% RH
Pressure Range	25 - 31 in Hg
Power Requirements	100V-120V, 220V-240V, 50/60 Hz
Analog Output Ranges	10V, 5V, 1V, 0.1V (selectable)
Recorder Offset	±10%
Included I/O	1 x Ethernet: 10/100Base-T 2 x RS232 (300-115,200 baud) 2 x USB device ports 8 x opto-isolated digital outputs 6 x opto-isolated digital inputs 4 x analog outputs
Optional I/O	1 x USB com port 1 x RS485 8 x analog inputs (0-10V, 12-bit) 4 x digital alarm outputs Multidrop RS232 3 x 4-20mA current outputs
Operating Temperature Range	5 - 40°C
Dimensions (HxWxD)	7" x 17" x 23.5" (178 x 432 x 597 mm)
Weight	28 lbs (12.7 kg)
Optional CO <sub>2</sub> Sensor Measurement Range	0-20% full scale CO <sub>2</sub>

Specifications subject to change without notice. All specifications are based on constant conditions.



family of monitoring instrumentation products, call us or visit our website at:

For more information about the Teledyne API



© 2018 Teledyne API Printed documents are uncontrolled. SAL000066C (DCN 7985) 10.01.18

